

Parts/Equipment Purchase Request Form



Name of equipment/Part **Thermocouple** Document Number:
Request No: Date: Nov-2025

Row	Category/Class	Description	QTY	Unit	Technical Specifications
1	Thermocouple	Type R, Rail Mounted, Measuring Range: 0-1200 °C, Thermowell connection and size: Flange RF #150 1 1/2 in, U Length: 700 mm, Thermowell Material: Inconel 601, Transmitter Brand: Endress+Hauser, Order Code: TMT72-AAP2AAAA1	51	Number	<u>Refer to The Attached Datasheets for Detailed Technical Specifications.</u>

Document checklist

Row	Description	Type	No	Yes	Doc No.	Page(s)	Technical Comment(s)
2	Attachments	Data Sheet	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Datasheet for Thermocouple	1	<u>The supplier is required to stamp and sign this file and all documents attached to this file.</u>
		Spec.FAT	<input type="checkbox"/>	<input type="checkbox"/>			
		Drawings	<input type="checkbox"/>	<input type="checkbox"/>			
		Image	<input type="checkbox"/>	<input type="checkbox"/>			
		ITP	<input type="checkbox"/>	<input checked="" type="checkbox"/>	INSPECTION TEST PLAN FOR THERMOCOUPLE	1	
		Name Plate	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Transmitter_pic	1	
		Standard	<input type="checkbox"/>	<input type="checkbox"/>			
		Catalogue	<input type="checkbox"/>	<input type="checkbox"/>			
		Other	<input type="checkbox"/>	<input type="checkbox"/>			

		Type of the document		Description	Type of the document		Description
3	Documents & Certificate	Calibration Certificate	<input checked="" type="checkbox"/>		NDT Certificate	<input type="checkbox"/>	
		Material Analysis	<input checked="" type="checkbox"/>		Certificate	<input checked="" type="checkbox"/>	
		Performance Test	<input checked="" type="checkbox"/>		Balance Certificate	<input type="checkbox"/>	
		Guarantee	<input checked="" type="checkbox"/>		Heat Treatment	<input type="checkbox"/>	
		Product Manual	<input checked="" type="checkbox"/>		Tensile Test	<input type="checkbox"/>	
		Sample Test	<input type="checkbox"/>		Hardness Test	<input type="checkbox"/>	
		Test Report	<input checked="" type="checkbox"/>		DataSheet	<input checked="" type="checkbox"/>	
		Drawings	<input type="checkbox"/>		Coating Certificate	<input type="checkbox"/>	
		QC Control Report	<input checked="" type="checkbox"/>			<input type="checkbox"/>	
Others: * Serial number must be mentioned in test reports and certificates, as they are essential for the QC procedure.							

4	Site Condition	Mean Sea Level: 1750 M Max. Recorded Temperature: +50°C Min. Recorded Temperature: -15°C Relative Humidity: 70%
---	----------------	--

Description :
The conditions that the bidder must comply are as follows:
 1. All test steps must be done with the presence of an inspector and based on the ITP certificate approved by this company. (The requested ITP certificate is attached to this file).
 2. The supplier must provide the above documents, including the Test Report, in the Final Book and submit them to this company after the final approval and test.
 3. Final QC procedure will be done at the buyer company.